

CERTIFICATE OF ANALYSIS

LOT NUMBER: 19DC0429
PRODUCT NAME: V-PURE
REFERENCE: V-PW-2000
EXPIRY DATE: 29.04.2022

TEST	SPECIFICATION	RESULT
ENDOTOXIN	< 0,25 EU/mL	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED