



VITROMED
Germany

VITROMED GmbH
Hans-Knöll-Str. 6
07745 Jena
GERMANY

Tel +49 3641 539 1976
Fax +49 3641 539 1977

support@vitromed.com
www.vitromed.com

CERTIFICATE OF ANALYSIS

LOT NUMBER: 20113VD25, 20079VD35, 20080VD40, 20081VD75, 20082VD20, 20083VD27, 20084VD30
PRODUCT NAME: V-DENUPET
REFERENCE: V-DEN-125, V-DEN-135, V-DEN-140, V-DEN-175, V-DEN-200, V-DEN-275, V-DEN-300
EXPIRY DATE: 10.02.2023
INDICATIONS FOR USE: V-DENUPET is used for the removal of cumulus oocyte complexes (COC) from oocyte prior to Intracytoplasmic Sperm Injection (ICSI) and In Vitro Fertilization (IVF) as well as for the handling of oocytes and embryos during assisted reproductive techniques (ART). V-DENUPET is not intended for biopsy of cells from oocytes or embryos.
STORAGE CONDITIONS: Keep away from sunlight and heat. Store at 15 – 25 °C. Keep dry.

Rx only

TEST	SPECIFICATION	RESULT
GAMMA STERILIZATION	25 - 45 kGy	PASSED
ENDOTOXIN	< 0.05 EU/device	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED