

## CERTIFICATE OF ANALYSIS

---

LOT NUMBER: 20154HO30, 20155HO35

PRODUCT NAME: V-ICSIPET HOLDING

REFERENCE: V-HOLD-30, V-HOLD-35

EXPIRY DATE: 06.04.2023

INDICATIONS FOR USE: V-ICSIPET-HOLDING is intended to use for fixing the oocyte or embryos in the implementation of an ICSI or biopsy.

STORAGE CONDITIONS: Keep away from sunlight and heat. Keep dry.

Rx only

| TEST                                      | SPECIFICATION    | RESULT |
|---|------------------|--------|
| GAMMA STERILIZATION                       | 25 - 45 kGy      | PASSED |
| ENDOTOXIN                                 | < 0.05 EU/device | PASSED |
| MEA 1-cell to expanded blastocyst in 96 h | ≥ 80 %           | PASSED |