



**VITROMED**  
Germany

VITROMED GmbH  
Hans-Knöll-Str. 6  
07745 Jena  
GERMANY

Tel +49 3641 539 1976  
Fax +49 3641 539 1977

support@vitromed.com  
www.vitromed.com

## CERTIFICATE OF ANALYSIS

---

LOT NUMBER: 20193VD25, 20194VD40, 20195VD50, 20196VD75, 20197VD20, 20198VD30

PRODUCT NAME: V-DENUPET

REFERENCE: V-DEN-125, V-DEN-140, V-DEN-150, V-DEN-175, V-DEN-200, V-DEN-300

EXPIRY DATE: 29.06.2023

INDICATIONS FOR USE: V-DENUPET is used for the removal of cumulus oocyte complexes (COC) from oocyte prior to Intracytoplasmic Sperm Injection (ICSI) and In Vitro Fertilization (IVF) as well as for the handling of oocytes and embryos during assisted reproductive techniques (ART). V-DENUPET is not intended for biopsy of cells from oocytes or embryos.

STORAGE CONDITIONS: Keep away from sunlight and heat. Store at 15 – 25 °C. Keep dry.

Rx only

TEST	SPECIFICATION	RESULT
GAMMA STERILIZATION	25 - 45 kGy	PASSED
ENDOTOXIN	< 0.05 EU/device	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED