

## CERTIFICATE OF ANALYSIS

---

LOT NUMBER: 20213HO30, 20214HO35

PRODUCT NAME: V-ICSIPET HOLDING

REFERENCE: V-HOLD-30, V-HOLD-35

EXPIRY DATE: 06.07.2023

INDICATIONS FOR USE: V-ICSIPET-HOLDING is intended to use for fixing the oocyte or embryos in the implementation of an ICSI or biopsy.

STORAGE CONDITIONS: Keep away from sunlight and heat. Keep dry.

Rx only

TEST	SPECIFICATION	RESULT
GAMMA STERILIZATION	25 - 45 kGy	PASSED
ENDOTOXIN	< 0.05 EU/device	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED