



VITROMED
Germany

VITROMED GmbH
Hans-Knöll-Str. 6
07745 Jena
GERMANY

Tel +49 3641 539 1976
Fax +49 3641 539 1977

support@vitromed.com
www.vitromed.com

CERTIFICATE OF ANALYSIS

LOT NUMBER: 20225HO30, 20245HO35

PRODUCT NAME: V-ICSIPET HOLDING

REFERENCE: V-HOLD-30, V-HOLD-35

EXPIRY DATE: 10.08.2023

INDICATIONS FOR USE: V-ICSIPET-HOLDING is intended to use for fixing the oocyte or embryos in the implementation of an ICSI or biopsy.

STORAGE CONDITIONS: Keep away from sunlight and heat. Keep dry.

Rx only

| TEST | SPECIFICATION | RESULT |
|---|------------------|--------|
| GAMMA STERILIZATION | 25 - 45 kGy | PASSED |
| ENDOTOXIN | < 0.05 EU/device | PASSED |
| MEA 1-cell to expanded blastocyst in 96 h | ≥ 80 % | PASSED |