

CERTIFICATE OF ANALYSIS

LOT NUMBER: 20276IN30V1

PRODUCT NAME: V-ICSIPET TESE V1 30°

REFERENCE: V-INJ-V1-30

EXPIRY DATE: 18.10.2023

INDICATIONS FOR USE: V-ICSIPET-INJECTION is intended to use for intracytoplasmic injection (ICSI) of a single sperm into the oocyte (with and without spike).

STORAGE CONDITIONS: Keep away from sunlight and heat. Keep dry.

Rx only

TEST	SPECIFICATION	RESULT
GAMMA STERILIZATION	25 - 45 kGy	PASSED
ENDOTOXIN	< 0.05 EU/device	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED