



VITROMED
Germany

VITROMED GmbH
Hans-Knöll-Str. 6
07745 Jena
GERMANY

Tel +49 3641 539 1976
Fax +49 3641 539 1977

support@vitromed.com
www.vitromed.com

CERTIFICATE OF ANALYSIS

LOT NUMBER: 21355HO20, 21356HO35
PRODUCT NAME: V-ICSIPET HOLDING
REFERENCE: V-HOLD-20, V-HOLD-35
EXPIRY DATE: 01.02.2024

INDICATIONS FOR USE: V-ICSIPET-HOLDING is intended to use for fixing the oocyte or embryos in the implementation of an ICSI or biopsy.

STORAGE CONDITIONS: Keep away from sunlight and heat. Store at 15 – 25 °C. Keep dry.

Rx only

TEST	SPECIFICATION	RESULT
GAMMA STERILIZATION	25 - 45 kGy	PASSED
ENDOTOXIN	< 0.05 EU/device	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED