

CERTIFICATE OF ANALYSIS

LOT NUMBER: 21357IN203, 21358IN353
PRODUCT NAME: V-ICSIPET INJECTION
REFERENCE: V-INJ-S3-20, V-INJ-S3-35
EXPIRY DATE: 01.02.2024

INDICATIONS FOR USE: V-ICSIPET-INJECTION is intended to use for intracytoplasmic injection (ICSI) of a single sperm into the oocyte (with and without spike).

STORAGE CONDITIONS: Keep away from sunlight and heat. Keep dry.

Rx only

TEST	SPECIFICATION	RESULT
GAMMA STERILIZATION	25 - 45 kGy	PASSED
ENDOTOXIN	< 0.05 EU/device	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED