

CERTIFICATE OF ANALYSIS

LOT NUMBER: 21373TR301, 21367TR351
PRODUCT NAME: V-ICSIPET TROPHECTODERM BIOPSY
REFERENCE: V-TEBP-V2-30, V-TEBP-V1-35
EXPIRY DATE: 01.02.2024

INDICATIONS FOR USE: V-ICSIPET-TROPHECTODERM BIOPSY is intended to use for the collection of trophoctoderm cells for subsequent preimplantation genetic diagnosis.

STORAGE CONDITIONS: Keep away from sunlight and heat. Store at 15 – 25 °C. Keep dry.

Rx only

TEST	SPECIFICATION	RESULT
GAMMA STERILIZATION	25 - 45 kGy	PASSED
ENDOTOXIN	< 0.05 EU/device	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED