

CERTIFICATE OF ANALYSIS

LOT NUMBER: 21390BL301
PRODUCT NAME: V-BLASTOMERE 30⁰
REFERENCE: V-BBP-V1-30
EXPIRY DATE: 08.03.2024

INDICATIONS FOR USE: V-ICSIPET-BLASTOMERE BIOPSY is intended to use for the collection of blastomeres for subsequent preimplantation genetic diagnosis.

STORAGE CONDITIONS: Keep away from sunlight and heat. Keep dry.

Rx only

TEST	SPECIFICATION	RESULT
GAMMA STERILIZATION	25 - 45 kGy	PASSED
ENDOTOXIN	< 0.05 EU/device	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED