



VITROMED
Germany

VITROMED GmbH
Wildenbruchstr. 15
07745 Jena
GERMANY

Tel +49 3641 539 1976
Fax +49 3641 539 1977

support@vitromed.com
www.vitromed.com

CERTIFICATE OF ANALYSIS

LOT NUMBER: 811M1402
PRODUCT NAME: V-ONESTEP Culture Medium
REFERENCE: V-OSM-20
EXPIRY DATE: 18.02.2019

| TEST | SPECIFICATION | RESULT |
|---|-----------------|--------|
| pH | 7,25-7,40 | 7,31 |
| OSMOLALITY | 257-273 mOsm/kg | 263 |
| BACTERIAL SCREEN | PASSED | PASSED |
| FUNGAL SCREEN | PASSED | PASSED |
| ENDOTOXIN | < 0,25 EU/mL | PASSED |
| Time Lapse MEA 1-cell to expanded blastocyst 120h | ≥ 80 % | PASSED |
| Blastocyst Cell Count | PASSED | PASSED |
| Color | Pink rose | PASSED |

i. A. Inena Shvornets