



VITROMED
Germany

VITROMED GmbH
Wildenbruchstr. 15
07745 Jena
GERMANY

Tel +49 3641 539 1976
Fax +49 3641 539 1977

support@vitromed.com
www.vitromed.com

CERTIFICATE OF ANALYSIS

LOT NUMBER: 812M0601
PRODUCT NAME: V-SPERM FREEZE
REFERENCE: V-SFM-12
EXPIRY DATE: 30.06.2020

TEST	SPECIFICATION	RESULT
pH	7,20-7,60	7,38
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
SPERM SURVIVAL TEST	≥ 80 %	PASSED
ENDOTOXIN	< 0,25 EU/mL	PASSED

i. A. Jrema Sherameti