



VITROMED
Germany

VITROMED GmbH
Hans-Knöll-Str. 6
07745 Jena
GERMANY

Tel +49 3641 539 1976
Fax +49 3641 539 1977

support@vitromed.com
www.vitromed.com

CERTIFICATE OF ANALYSIS

LOT NUMBER: 902M0401

PRODUCT NAME: V-GRAD Kit

REFERENCE: V-GRD-KIT

EXPIRY DATE: 31.08.2020

| TEST | SPECIFICATION | RESULT |
|------------------------|-------------------|--------|
| pH (V-GRAD 40) | 7,20-7,60 | 7,40 |
| pH (V-GRAD 80) | 7,20-7,60 | 7,35 |
| OSMOLALITY (V-GRAD 40) | 310 - 340 mOsm/kg | 328 |
| OSMOLALITY (V-GRAD 80) | 310-330 mOsm/kg | 323 |
| ENDOTOXIN | < 0,5 EU/mL | PASSED |
| BACTERIAL SCREEN | PASSED | PASSED |
| FUNGAL SCREEN | PASSED | PASSED |
| SPERM SURVIVAL TEST | ≥ 80 % | PASSED |