



**VITROMED**  
Germany

VITROMED GmbH  
Hans-Knöll-Str. 6  
07745 Jena  
GERMANY

Tel +49 3641 539 1976  
Fax +49 3641 539 1977

support@vitromed.com  
www.vitromed.com

## CERTIFICATE OF ANALYSIS

---

LOT NUMBER: 902M3401  
PRODUCT NAME: V-PVP 7%  
REFERENCE: V-PVP-7  
EXPIRY DATE: 29.02.2020

TEST	SPECIFICATION	RESULT
pH	7,25-7,45	7,31
OSMOLALITY	300-330 mOsm/kg	319
ENDOTOXIN	< 0,5 EU/mL	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED
FUNGAL SCREEN	PASSED	PASSED
BACTERIAL SCREEN	PASSED	PASSED