



**VITROMED**  
Germany

VITROMED GmbH  
Hans-Knöll-Str. 6  
07745 Jena  
GERMANY

Tel +49 3641 539 1976  
Fax +49 3641 539 1977

support@vitromed.com  
www.vitromed.com

## CERTIFICATE OF ANALYSIS

---

LOT NUMBER: 904M0501  
PRODUCT NAME: V-SPERM WASH  
REFERENCE: V-SWM-100  
EXPIRY DATE: 30.04.2020

| TEST                                      | SPECIFICATION   | RESULT |
|---|-----------------|--------|
| pH  | 7,00-7,40       | 7,31   |
| OSMOLALITY                                | 257-273 mOsm/kg | 266    |
| BACTERIAL SCREEN                          | PASSED          | PASSED |
| FUNGAL SCREEN                             | PASSED          | PASSED |
| ENDOTOXIN                                 | < 0,25 EU/mL    | PASSED |
| MEA 1-cell to expanded blastocyst in 96 h | ≥ 80 %          | PASSED |