



**VITROMED**  
Germany

VITROMED GmbH  
Hans-Knöll-Str. 6  
07745 Jena  
GERMANY

Tel +49 3641 539 1976  
Fax +49 3641 539 1977

support@vitromed.com  
www.vitromed.com

## CERTIFICATE OF ANALYSIS

---

LOT NUMBER: 904M2301  
PRODUCT NAME: V-HEPES PLUS 100  
REFERENCE: V-HTF-P1  
EXPIRY DATE: 30.04.2020

TEST	SPECIFICATION	RESULT
pH	7,00-7,40	7,32
OSMOLALITY	257-273 mOsm/kg	268
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0,25 EU/mL	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED