



VITROMED
Germany

VITROMED GmbH
Hans-Knöll-Str. 6
07745 Jena
GERMANY

Tel +49 3641 539 1976
Fax +49 3641 539 1977

support@vitromed.com
www.vitromed.com

CERTIFICATE OF ANALYSIS

LOT NUMBER: 904M6101
PRODUCT NAME: V-OIL 500 P / OEM
REFERENCE: V-OIL-P500 / OEM
EXPIRY DATE: 31.10.2020

TEST	SPECIFICATION	RESULT
Peroxide value (POV) determination	< 0,1 mEq/kG	PASSED
ENDOTOXIN	< 0,1 EU/ mL	PASSED
STERILITY by sterile filtration over 0,2 µm filter	PASSED	PASSED
Time Lapse MEA 1-cell to expanded blastocyst 120h	≥ 80 %	PASSED
Blastocyst Cell Count	PASSED	PASSED