



VITROMED
Germany

VITROMED GmbH
Hans-Knöll-Str. 6
07745 Jena
GERMANY

Tel +49 3641 539 1976
Fax +49 3641 539 1977

support@vitromed.com
www.vitromed.com

CERTIFICATE OF ANALYSIS

LOT NUMBER: 905M01G01
PRODUCT NAME: V-GRD-100
REFERENCE: V-GRD00_100G
EXPIRY DATE: 30.11.2020

TEST	SPECIFICATION	RESULT
pH	7,20-7,60	7,35
OSMOLALITY	300-330 mOsm/kg	325
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0,5 EU/mL	PASSED
SPERM SURVIVAL TEST	≥ 80 %	PASSED