

## CERTIFICATE OF ANALYSIS

---

LOT NUMBER: 905M02G02  
PRODUCT NAME: V-GRD-80  
REFERENCE: V-GRD83\_100G  
EXPIRY DATE: 30.11.2020

TEST	SPECIFICATION	RESULT
pH	7,20-7,60	7,25
OSMOLALITY	310-330 mOsm/kg	329
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0,5 EU/mL	PASSED
SPERM SURVIVAL TEST	≥ 80 %	PASSED