

CERTIFICATE OF ANALYSIS

LOT NUMBER: 905M03G02
PRODUCT NAME: V-GRAD 40%
REFERENCE: V-GRD43_100G
EXPIRY DATE: 30.11.2020

TEST	SPECIFICATION	RESULT
pH	7,20-7,60	7,34
OSMOLALITY	310 - 340 mOsm/kg	331
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0,5 EU/mL	PASSED
SPERM SURVIVAL TEST	≥ 80 %	PASSED