

## CERTIFICATE OF ANALYSIS

---

LOT NUMBER: 906M0901  
PRODUCT NAME: V-SPERM WASH KIT  
REFERENCE: V-SWM-02  
EXPIRY DATE: 30.06.2020

TEST	SPECIFICATION	RESULT
pH	7,00-7,40	7,32
OSMOLALITY	257-273 mOsm/kg	265
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0,25 EU/mL	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED