

CERTIFICATE OF ANALYSIS

LOT NUMBER: 906M1402
PRODUCT NAME: V-ONESTEP Culture Medium
REFERENCE: V-OSM-20
EXPIRY DATE: 18.09.2019

TEST	SPECIFICATION	RESULT
pH	7,25-7,40	7,36
OSMOLALITY	257-273 mOsm/kg	265
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0,25 EU/mL	PASSED
Time Lapse MEA 1-cell to expanded blastocyst 120h	≥ 80 %	PASSED
Blastocyst Cell Count	PASSED	PASSED
Color	Pink rose	PASSED