

## CERTIFICATE OF ANALYSIS

---

LOT NUMBER: 907M0601  
PRODUCT NAME: V-SPERM FREEZE  
REFERENCE: V-SFM-12  
EXPIRY DATE: 31.01.2021

TEST	SPECIFICATION	RESULT
pH	7,20-7,60	7,54
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
SPERM SURVIVAL TEST	≥ 80 %	PASSED
ENDOTOXIN	< 0,25 EU/mL	PASSED