

CERTIFICATE OF ANALYSIS

LOT NUMBER: 907M3101
PRODUCT NAME: V-OIL 500 P CE
REFERENCE: V-OIL-P500
EXPIRY DATE: 31.01.2021

TEST	SPECIFICATION	RESULT
Peroxide value (POV) determination	< 0,1 mEq/kg	PASSED
ENDOTOXIN	< 0,1 EU/ mL	PASSED
STERILITY by sterile filtration over 0,2 µm filter	PASSED	PASSED
Blastocyst Cell Count	PASSED	PASSED
Time Lapse MEA 1-cell to expanded blastocyst 120h	≥ 80 %	PASSED