

## CERTIFICATE OF ANALYSIS

---

LOT NUMBER: 908M1301  
PRODUCT NAME: V-ONESTEP Culture Medium  
REFERENCE: V-OSM-10  
EXPIRY DATE: 12.11.2019

TEST	SPECIFICATION	RESULT
pH	7,25-7,40	7,40
OSMOLALITY	257-273 mOsm/kg	262
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0,25 EU/mL	PASSED
Time Lapse MEA 1-cell to expanded blastocyst 120h	≥ 80 %	PASSED
Blastocyst Cell Count	PASSED	PASSED
Color	Pink rose	PASSED