



**VITROMED**  
Germany

VITROMED GmbH  
Hans-Knöll-Str. 6  
07745 Jena  
GERMANY

Tel +49 3641 539 1976  
Fax +49 3641 539 1977

support@vitromed.com  
www.vitromed.com

## CERTIFICATE OF ANALYSIS

---

LOT NUMBER: 908M1403  
PRODUCT NAME: V-ONESTEP Culture Medium  
REFERENCE: V-OSM-20  
EXPIRY DATE: 26.11.2019

| TEST  | SPECIFICATION   | RESULT |
|---|-----------------|--------|
| pH  | 7,25-7,40       | 7,38   |
| OSMOLALITY  | 257-273 mOsm/kg | 265    |
| BACTERIAL SCREEN                                  | PASSED          | PASSED |
| FUNGAL SCREEN                                     | PASSED          | PASSED |
| ENDOTOXIN   | < 0,25 EU/mL    | PASSED |
| Time Lapse MEA 1-cell to expanded blastocyst 120h | ≥ 80 %          | PASSED |
| Blastocyst Cell Count                             | PASSED          | PASSED |
| Color   | Pink rose       | PASSED |