

CERTIFICATE OF ANALYSIS

LOT NUMBER: 908M2301
PRODUCT NAME: V-HEPES PLUS 100
REFERENCE: V-HTF-P1
EXPIRY DATE: 31.08.2020

TEST	SPECIFICATION	RESULT
pH	7,00-7,40	7,40
OSMOLALITY	257-273 mOsm/kg	262
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0,25 EU/mL	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED