

CERTIFICATE OF ANALYSIS

LOT NUMBER: 908M3501
PRODUCT NAME: V-PVP 10%
REFERENCE: V-PVP-10
EXPIRY DATE: 31.08.2020

TEST	SPECIFICATION	RESULT
pH	7,25-7,45	7,36
OSMOLALITY	300-330 mOsm/kg	316
ENDOTOXIN	< 0,5 EU/mL	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED
FUNGAL SCREEN	PASSED	PASSED
BACTERIAL SCREEN	PASSED	PASSED