

## CERTIFICATE OF ANALYSIS

---

LOT NUMBER: 908M3601  
PRODUCT NAME: V-HYADASE  
REFERENCE: V-HYA  
EXPIRY DATE: 31.08.2020

TEST	SPECIFICATION	RESULT
pH	7,20-7,40	7,30
OSMOLALITY	257-273 mOsm/kg	268
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 1,00 EU/mL	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED