



VITROMED
Germany

VITROMED GmbH
Hans-Knöll-Str. 6
07745 Jena
GERMANY

Tel +49 3641 539 1976
Fax +49 3641 539 1977

support@vitromed.com
www.vitromed.com

CERTIFICATE OF ANALYSIS

LOT NUMBER: 909M0101
PRODUCT NAME: V-GRAD 100%
REFERENCE: V-GRD-100
EXPIRY DATE: 31.03.2021

TEST	SPECIFICATION	RESULT
pH	7,20-7,60	7,33
OSMOLALITY	300-330 mOsm/kg	325
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0,5 EU/mL	PASSED
SPERM SURVIVAL TEST	≥ 80 %	PASSED