

CERTIFICATE OF ANALYSIS

LOT NUMBER: 909M0201
PRODUCT NAME: V-GRAD 80%
REFERENCE: V-GRD-80
EXPIRY DATE: 31.03.2021

TEST	SPECIFICATION	RESULT
pH	7,20-7,60	7,38
OSMOLALITY	310-330 mOsm/kg	326
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0,5 EU/mL	PASSED
SPERM SURVIVAL TEST	≥ 80 %	PASSED