



VITROMED
Germany

VITROMED GmbH
Hans-Knöll-Str. 6
07745 Jena
GERMANY

Tel +49 3641 539 1976
Fax +49 3641 539 1977

support@vitromed.com
www.vitromed.com

CERTIFICATE OF ANALYSIS

LOT NUMBER: 909M0301
PRODUCT NAME: V-GRAD 40%
REFERENCE: V-GRD-40
EXPIRY DATE: 31.03.2021

TEST	SPECIFICATION	RESULT
pH	7,20-7,60	7,41
OSMOLALITY	310 - 340 mOsm/kg	337
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0,5 EU/mL	PASSED
SPERM SURVIVAL TEST	≥ 80 %	PASSED