

CERTIFICATE OF ANALYSIS

LOT NUMBER: 910M3601
PRODUCT NAME: V-HYADASE
REFERENCE: V-HYA
EXPIRY DATE: 31.10.2020

TEST	SPECIFICATION	RESULT
pH	7,20-7,40	7,399
OSMOLALITY	257-273 mOsm/kg	261
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 1,00 EU/mL	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED