

CERTIFICATE OF ANALYSIS

LOT NUMBER: 911K4K401
PRODUCT NAME: V-GRAD Kit
REFERENCE: V-GRD-KIT
EXPIRY DATE: 31.05.2021

TEST	SPECIFICATION	RESULT
pH (V-GRAD 40)	7.20-7.60	7,40
pH (V-GRAD 80)	7.20-7.60	7,32
OSMOLALITY (V-GRAD 40)	310 - 340 mOsm/kg	328
OSMOLALITY (V-GRAD 80)	310-330 mOsm/kg	325
ENDOTOXIN	< 0.5 EU/mL	PASSED
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
SPERM SURVIVAL TEST	≥ 80 %	PASSED
SARS-CoV-2 / COVID-19 PCR detection	PASSED	PASSED