

CERTIFICATE OF ANALYSIS

LOT NUMBER: 911M0301
PRODUCT NAME: V-GRAD 40%
REFERENCE: V-GRD-40
EXPIRY DATE: 31.05.2021

TEST	SPECIFICATION	RESULT
pH	7.20-7.60	7,34
OSMOLALITY	310 - 340 mOsm/kg	327
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0.5 EU/mL	PASSED
SPERM SURVIVAL TEST	≥ 80 %	PASSED
SARS-CoV-2 / COVID-19 PCR detection	PASSED	PASSED