

CERTIFICATE OF ANALYSIS

LOT NUMBER: 911M0601
PRODUCT NAME: V-SPERM FREEZE
REFERENCE: V-SFM-12
EXPIRY DATE: 31.05.2021

TEST	SPECIFICATION	RESULT
pH	7,20-7,60	7,46
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
SPERM SURVIVAL TEST	≥ 80 %	PASSED
ENDOTOXIN	< 0,25 EU/mL	PASSED