

CERTIFICATE OF ANALYSIS

LOT NUMBER: 911M1401

PRODUCT NAME: V-ONESTEP Culture Medium

REFERENCE: V-OSM-20

EXPIRY DATE: 04.02.2020

TEST	SPECIFICATION	RESULT
pH	7,25-7,40	7,299
OSMOLALITY	257-273 mOsm/kg	263
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0,25 EU/mL	PASSED
Time Lapse MEA 1-cell to expanded blastocyst 120h	≥ 80 %	PASSED
Blastocyst Cell Count	PASSED	PASSED
Color	Pink rose	PASSED