

CERTIFICATE OF ANALYSIS

LOT NUMBER: 912M1301
PRODUCT NAME: V-ONESTEP Culture Medium
REFERENCE: V-OSM-10
EXPIRY DATE: 03.03.2020

TEST	SPECIFICATION	RESULT
pH	7.25-7.40	7.365
OSMOLALITY	257-273 mOsm/kg	263
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0.25 EU/mL	PASSED
Time Lapse MEA 1-cell to expanded blastocyst 120h	≥ 80 %	PASSED
Blastocyst Cell Count	PASSED	PASSED
Color	Pink rose	PASSED