

CERTIFICATE OF ANALYSIS

LOT NUMBER: H87209-1
PRODUCT NAME: V-TRANSIN STERILE LUER
REFERENCE: V-TRI-ET
EXPIRY DATE: 30.04.2024

TEST	SPECIFICATION	RESULT
ENDOTOXIN	< 0,05 EU/mL	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED