

## CERTIFICATE OF ANALYSIS

---

LOT NUMBER: H88748-1  
PRODUCT NAME: V-TRANSIN  
REFERENCE: V-TRI\_ET  
EXPIRY DATE: 16.02.2025

TEST	SPECIFICATION	RESULT
ENDOTOXIN	< 0.05 EU/device	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED