

## CERTIFICATE OF ANALYSIS

---

LOT NUMBER: N005705, N005710, N005706, N005707, N005708, N005709  
PRODUCT NAME: V-DENUPET-SP  
REFERENCE: V-DEN-SP125, V-DEN-SP135, V-DEN-SP140, V-DEN-SP150, V-DEN-SP175, V-DEN-SP275  
EXPIRY DATE: 28.01.2024

TEST	SPECIFICATION	RESULT
GAMMA STERILIZATION	25 - 45 kGy	PASSED
ENDOTOXIN	< 0.05 EU/device	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED