



VITROMED
Germany

VITROMED GmbH
Raiffeisenstr 15a
40764 Langenfeld
GERMANY

Tel +49 2173-20041-30

www.vitromed.com

CERTIFICATE OF ANALYSIS

LOT NUMBER: 102VM2501
PRODUCT NAME: V-HEPES PLUS (500 mL)
REFERENCE: V-HTF-P5
EXPIRY DATE: 24-Feb-2022

TEST	SPECIFICATION	RESULT
pH	7.00-7.40	7.27
OSMOLALITY	257-273 mOsm/kg	263
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0.25 EU/mL	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED
Color	PASSED	PASSED