

CERTIFICATE OF ANALYSIS

LOT NUMBER: 001M6301
PRODUCT NAME: V-OIL 500 G
REFERENCE: V-OIL-G500
EXPIRY DATE: 31.07.2021

| TEST | SPECIFICATION | RESULT |
|--|---------------|--------|
| Peroxide value (POV) determination | < 0.1 mEq/kg | PASSED |
| ENDOTOXIN | < 0.1 EU/ mL | PASSED |
| STERILITY by sterile filtration over 0,2 µm filter | PASSED | PASSED |
| Blastocyst Cell Count | PASSED | PASSED |
| Time Lapse MEA 1-cell to expanded blastocyst 120h | ≥ 80 % | PASSED |

