

## CERTIFICATE OF ANALYSIS

---

LOT NUMBER: 004VM0501  
PRODUCT NAME: V-SPERM WASH  
REFERENCE: V-SWM-100  
EXPIRY DATE: 21.04.2021

TEST	SPECIFICATION	RESULT
pH	7.00-7.40	7,26
OSMOLALITY	257-273 mOsm/kg	263
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0.25 EU/mL	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED
SARS-CoV-2 / COVID-19 PCR detection	PASSED	PASSED

