



**VITROMED**  
Germany

VITROMED GmbH  
Raiffeisenstraße 15a  
40764 Langenfeld  
GERMANY

Tel +49 2173-20041-30

www.vitromed.com

## CERTIFICATE OF ANALYSIS

---

LOT NUMBER: 006M1401  
PRODUCT NAME: V-ONESTEP Culture Medium  
REFERENCE: V-OSM-20  
EXPIRY DATE: 01-Sep-2020

TEST	SPECIFICATION	RESULT
pH	7.25 – 7.40	7.36
OSMOLALITY	257 - 273	263
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	<0,25 EU/mL	PASSED
Time Lapse MEA 1-cell to expanded blastocysts 120h	≥ 80%	PASSED
Blastocyst Cell Count	PASSED	PASSED
Color	Pink rose	PASSED
SARS-CoV-2 / COVID-19 PCR detection	PASSED	PASSED

*Lela Abazi*