

## CERTIFICATE OF ANALYSIS

---

LOT NUMBER: 007M1301  
PRODUCT NAME: V-ONESTEP Culture Medium  
REFERENCE: V-OSM-10  
EXPIRY DATE: 06-Oct-2020

| TEST   | SPECIFICATION | RESULT |
|--|---------------|--------|
| pH   | 7.25 – 7.40   | 7.36   |
| OSMOLALITY   | 257 - 273     | 265    |
| BACTERIAL SCREEN                                   | PASSED        | PASSED |
| FUNGAL SCREEN                                      | PASSED        | PASSED |
| ENDOTOXIN  | <0,25 EU/mL   | PASSED |
| Time Lapse MEA 1-cell to expanded blastocysts 120h | ≥ 80%         | PASSED |
| Blastocyst Cell Count                              | PASSED        | PASSED |
| Color  | Pink rose     | PASSED |
| SARS-CoV-2 / COVID-19 PCR detection                | PASSED        | PASSED |