

CERTIFICATE OF ANALYSIS

LOT NUMBER: 007M1402
PRODUCT NAME: V-ONESTEP Culture Medium
REFERENCE: V-OSM-20
EXPIRY DATE: 25-Oct-2020

TEST	SPECIFICATION	RESULT
pH	7.25 – 7.40	7.37
OSMOLALITY	257 - 273	267
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	<0,25 EU/mL	PASSED
Time Lapse MEA 1-cell to expanded blastocysts 120h	≥ 80%	PASSED
Blastocyst Cell Count	PASSED	PASSED
Color	Pink rose	PASSED
SARS-CoV-2 / COVID-19 PCR detection	PASSED	PASSED