



VITROMED  
Germany

VITROMED GmbH  
Raiffeisenstr 15a  
40764 Langenfeld  
GERMANY

Tel +49 2173-20041-30

[www.vitromed.com](http://www.vitromed.com)

## CERTIFICATE OF ANALYSIS

---

LOT NUMBER: 008VM2101

PRODUCT NAME: V-HEPES

REFERENCE: V-HTF-1

EXPIRY DATE: 19.08.2021

TEST	SPECIFICATION	RESULT
pH	7.00-7.40	7.24
OSMOLALITY	257-273 mOsm/kg	259
BACTERIAL SCREEN	PASSED	PASSED
FUNGAL SCREEN	PASSED	PASSED
ENDOTOXIN	< 0.25 EU/mL	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED
Color	PASSED	PASSED