

## CERTIFICATE OF ANALYSIS

---

LOT NUMBER: 009VM0001

PRODUCT NAME: V-PURE

REFERENCE: V-PW-1000

EXPIRY DATE: 2023-09-22

TEST	SPECIFICATION	RESULT
ENDOTOXIN	< 0.05 EU/mL	PASSED
MEA 1-cell to expanded blastocyst in 96 h	≥ 80 %	PASSED

